

Declaration of Conformity to MDR 2017/745 concerning Medical Devices

Manufacturer:	Beijing Choice Electronic Technology Co., Ltd. 2nd Floor, 3rd Floor and Room 410-412 4th Floor, No. 2 Building, No. 9 Shuangyuan Road, Shijingshan District, 100041 Beijing, PEOPLE'S REPUBLIC OF CHINA SRN: CN-MF-000005776 Tel: +86-10-88203956 Fax: +86-10-88204632 E-mail: cc@choicemmed.com
European Representative:	Shanghai International Holding Corp. GmbH (Europe) Eiffestrasse 80, 20537 Hamburg, Germany Tel: +49-40-2513175 Fax: +49-40-255726 E-mail: shholding@hotmail.com
Product:	Pulse Oximeter
Model:	MD300CN130, MD300CN140, MD300CN150, MD300CN160 MD300C5, MD300C52, MD300C53, MD300C54, MD300C55 MD300CN310, MD300CN320, MD300CN330, MD300CN340, MD300CN350, MD300CN360, MD300CN390, MD300CN356, PX-101
Basic UDI-DI:	69428204SPO2VG
UMDNS Code:	17148
Classification:	Class IIa, rule 10 to Annex VIII of MDR 2017/745
Conformity assessment Route:	Annex IX chapter I, chapter III, and section.4

**We, the manufacturer, herewith declare that the stated medical devices
meet the transposition into national law, the provisions of Regulation (EU) 2017/745
of the European Parliament and of the Council of 5 April 2017 on medical devices.**

All supporting documentation is retained at the premises of the manufacturer.

We, the manufacturer, are sole responsible for the DoC.

Standards applied:

EN ISO 13485:2016/A11:2021 Medical devices- Quality management systems- Requirements for regulatory purposes

EN ISO 14971:2019/A11:2021 Medical devices - Application of risk management to medical devices

IEC 60601-1: 2005+A1:2012+A2:2020 Medical electrical equipment-Part 1: General requirements for safety

IEC 60601-1-2: 2020 Medical electrical equipment Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests

IEC 60601-1-11: 2015+A1:2020 Medical electrical equipment. Part 1-11: General requirements for basic safety and essential performance. Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment

EN ISO 80601-2-61:2019 Medical electrical equipment —Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment

EN 60601-1-6:2010+A2-2021 Medical electrical equipment -- Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability

EN 62304-2006+A1-2015 Medical device software-Software life-cycle processes

EN ISO 10993-1-2020 Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process

EN ISO 10993-5:2009 Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity

ISO 10993-10:2021 Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization

EN ISO 20417:2021 Medical devices - Information supplied by the manufacture

EN ISO 15223-1:2021 Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements

Notified Body: TÜV SÜD Product service GmbH
Ridlerstr 65, D-80339 München, Germany


Identification Number: 0123

(EC) Certificate(s): No. G10 057571 0008 Rev.00

Start of CE-marking: 2023-03-06

Place, Date of Declaration: Beijing, 2023-03-06

Signature:



Name: Haiying Zhao

Position: Quality Director